

ODOR	
Test Description	Determination of odor in water.
Test Use	Useful for evaluating drinking waters.
Test Department	Inorganic Chemistry: Phone 860-920-6666/6667 Fax 860-920-6670
Methodology	Sensory Test
Availability	Year-round
Sample Requirements	One (1) sample collected in a pre-cleaned plastic bottle
Container type /Preservative	½ Gallon plastic Chem Bottle. No preservative.
Collection Instructions (Note 1)	For taps, remove aerators and let water run 4-5 minutes. For outdoor locations, sampling location should be in accordance with a preapproved quality assurance project plan.
Sample Holding Time & Transport	Samples are iced or refrigerated and kept at 4°±2°C from time of collection until analysis. Samples must be analyzed within 24 hours of collection.
Unacceptable Conditions	Incomplete requisition form. Insufficient sample volume. Samples received beyond the 24-hour holding time.
Requisition Form	Use the Inorganic Chemistry Drinking Water form.
Required Information	Fill out entire requisition form.
Limitations	
Additional Comments	

Note 1: See *New England States Environmental Sampling Guide*, latest edition.

<https://www.epa.gov/sites/production/files/2015-06/documents/NE-States-Sample-Collection-Manual.pdf>